

MRB Workshop Services
Maryland House
23 Rutland Street
Swinton
Manchester
M27 6AU
Tel: 0161 793 4040
Fax: 0161 822 0070



Total project
management.....
Total workshop
solution.....

CERTIFICATE OF CALIBRATION

The listed equipment has been tested. Calibrated and has performed in accordance with the CUSTOMERS SPECIFICATION. Each of the readings listed are within tolerance laid down unless indicated otherwise

Customer details: Preston Scania Preston	Site Address:	Report No	Date Of Calibration
		MRB2019-09-1361	17/09/2019
			Page 22
		Previous Calibration date	Next Calibration Due
		UNKNOWN	17/09/2020
Description and Identification of the Equipment:-			
I.D. No	312088	Description	Torque wrench
Serial No		Manu	Norbar
Cal, Freq,	12 Months	Location	
		Model Details	
		Operation Range	200-750 lbft
		Customer Spec	+/- 4 %

Calibration Details	Traceability Information
TEST 1 - Cust 750 lbft TEST 1 - Master 747.2 lbft TEST 1 - Result PASS TEST 2 - Cust 500 lbft TEST 2 - Actual 514.1 lbft TEST 2 - Result PASS TEST 3 - Cust 200 lbft TEST 3 - Actual 197.7 lbft TEST 3 - Result PASS	Below are details of the master certificate:- equipment which has been calibrated at a UKAS laboratory to a specification traceable to national standards. Below are details of the master certificate:- ASSET CODE: 17475 EQUIP TYPE: Torque wrench analyser MANU: Norbar Torque Tool LTD CAL CERT No: 237273 UKAS LAB No: 0256 LAST CAL: 05/04/2019 NEXT CAL DUE: 05/04/2020
Additional comments/observations:-	CALIBRATION RESULT PASS

Calibration carried out by	Calibration performed by Engineer	Signatory for Calibration & Consultancy Service
MRB Workshop services Ltd Maryland House 23 Rutland street Swinton Manchester, M27 6AU	Name A.LEES Position Engineer Sign A.LEES	Name M.BATES Date 21/10/2019 Signature

