

MRB Workshop Services
 Maryland House
 23 Rutland Street
 Swinton
 Manchester
 M27 6AU
 Tel: 0161 793 4040
 Fax: 0161 822 0070



Total project management.....
 Total workshop solution.....

CERTIFICATE OF CALIBRATION

The listed equipment has been tested. Calibrated and has performed in accordance with the CUSTOMERS SPECIFICATION. Each of the readings listed are within tolerance laid down unless indicated otherwise

Customer details: Preston Scania	Site Address:	Report No MRB2019-09-1368	Date Of Calibration 17/09/2019
Preston			Page 29
		Previous Calibration date UNKNOWN	Next Calibration Due 17/09/2020
Description and Identification of the Equipment:-			
I.D. No	214803458	Description	Torque wrench
Serial No		Manu	Snap on
Cal, Freq,	6 Months	Location	
		Model Details	Operation Range 7-135 nm
			Customer Spec +/- 4 %

Calibration Details	Traceability Information
TEST 1 - Cust 135 nm TEST 1 - Master 131.1 nm TEST 1 - Result PASS TEST 2 - Cust 70 nm TEST 2 - Actual 67.5 nm TEST 2 - Result PASS TEST 3 - Cust 10 nm TEST 3 - Actual 9.7 nm TEST 3 - Result PASS	Below are details of the master certificate:- equipment which has been calibrated at a UKAS laboratory to a specification traceable to national standards. Below are details of the master certificate:- ASSET CODE: 17475 EQUIP TYPE: Torque wrench analyser MANU: Norbar Torque Tool LTD CAL CERT No: 237273 UKAS LAB No: 0256 LAST CAL: 05/04/2019 NEXT CAL DUE: 05/04/2020
Additional comments/observations:-	CALIBRATION RESULT Pass

Calibration carried out by	Calibration performed by Engineer	Signatory for Calibration & Consultancy Service	
MRB Workshop services Ltd Maryland House 23 Rutland street Swinton Manchester, M27 6AU	Name A.LEES Position Engineer Sign A.LEES	Name M.BATES Date 21/10/2019 Signature	

