

MRB Workshop Services  
Maryland House  
23 Rutland Street  
Swinton  
Manchester  
M27 6AU  
Tel: 0161 793 4040  
Fax: 0161 822 0070



Total project  
management.....  
Total workshop  
solution.....

## CERTIFICATE OF CALIBRATION

The listed equipment has been tested. Calibrated and has performed in accordance with the CUSTOMERS SPECIFICATION. Each of the readings listed are within tolerance laid down unless indicated otherwise

<b>Customer details:</b> Preston Scania  Preston	<b>Site Address:</b>	<b>Report No</b>	<b>Date Of Calibration</b>
		MRB2019-09-1376	17/09/2019
			<b>Page</b> 37
		<b>Previous Calibration date</b>	<b>Next Calibration Due</b>
		UNKNOWN	17/09/2020
<b>Description and Identification of the Equipment:-</b>			
<b>I.D. No</b>	922081	<b>Description</b>	Torque wrench
<b>Serial No</b>		<b>Manu</b>	Norbar
<b>Cal, Freq,</b>	6 Months	<b>Location</b>	
		<b>Model Details</b>	operation range 200-750 lb ft
		<b>Customer Spec</b>	+/- 4 %

<b>Calibration Details</b>	<b>Traceability Information</b>
TEST 1 - Cust 750 lb ft TEST 1 - Master 767.9 lb ft TEST 1 - Result PASS  TEST 2 - Cust 500 lb ft TEST 2 - Actual 498.2 lb ft TEST 2 - Result PASS  TEST 3 - Cust 250 lb ft TEST 3 - Actual 251.4 lb ft TEST 3 - Result PASS	Below are details of the master certificate:- equipment which has been calibrated at a UKAS laboratory to a specification traceable to national standards. Below are details of the master certificate:-  ASSET CODE: 17475 EQUIP TYPE: Torque wrench analyser MANU: Norbar Torque Tool LTD CAL CERT No: 237273 UKAS LAB No: 0256 LAST CAL: 05/04/2019 NEXT CAL DUE: 05/04/2020
<b>Additional comments/observations:-</b>	<b>CALIBRATION RESULT</b> Pass

<b>Calibration carried out by</b>	<b>Calibration performed by Engineer</b>	<b>Signatory for Calibration &amp; Consultancy Service</b>
MRB Workshop services Ltd Maryland House 23 Rutland street Swinton Manchester, M27 6AU	Name A.LEES Position Engineer  Sign A.LEES	Name M.BATES Date 21/10/2019  Signature

