

MRB Workshop Services
 Maryland House
 23 Rutland Street
 Swinton
 Manchester
 M27 6AU
 Tel: 0161 793 4040
 Fax: 0161 822 0070



Total project management.....
 Total workshop solution.....

CERTIFICATE OF CALIBRATION

The listed equipment has been tested. Calibrated and has performed in accordance with the CUSTOMERS SPECIFICATION. Each of the readings listed are within tolerance laid down unless indicated otherwise

Customer details: Haydock Commercials	Site Address:	Report No MRB2019-10-1398	Date Of Calibration 21/10/2019
Omega M62			Page 51
		Previous Calibration date UNKNOWN	Next Calibration Due 20/10/2020
Description and Identification of the Equipment:-			
I.D. No	B019531	Description	Torque Wrench
Serial No		Manu	Britool
Cal, Freq,	6 Monthly	Location	JH t/box
		Model Details	operation range 20-100lbf/ft
			Customer Spec +/- 4%

Calibration Details	Traceability Information
TEST 1 - Cust TEST 1 - Master TEST 1 - Result PASS	98.3 100 Below are details of the master certificate:- equipment which has been calibrated at a UKAS laboratory to a specification traceable to national standards. Below are details of the master certificate:-
TEST 2 - Cust TEST 2 - Actual TEST 2 - Result PASS	59 60 ASSET CODE: 17475 EQUIP TYPE: Torque Wrench Analyser MANU: Norbar Torque Tools LTD
TEST 3 - Cust TEST 3 - Actual TEST 3 - Result PASS	19.6 20 CAL CERT No: 237273 UKAS LAB No: 0256 LAST CAL: 09/04/2019 NEXT CAL DUE: 09/04/2020
Additional comments/observations:- Adjust into tolerance	CALIBRATION RESULT Pass

Calibration carried out by	Calibration performed by Engineer	Signatory for Calibration & Consultancy Service	
MRB Workshop services Ltd Maryland House 23 Rutland street Swinton Manchester, M27 6AU	Name A.LEES Position Engineer Sign A.LEES	Name M.BATES Date 21/10/2019 Signature	

