

MRB Workshop Services
 Maryland House
 23 Rutland Street
 Swinton
 Manchester
 M27 6AU
 Tel: 0161 793 4040
 Fax: 0161 822 0070



Total project
 management.....
 Total workshop
 solution.....

CERTIFICATE OF CALIBRATION

The listed equipment has been tested. Calibrated and has performed in accordance with the CUSTOMERS SPECIFICATION. Each of the readings listed are within tolerance laid down unless indicated otherwise

Customer details: Haydock Commercials	Site Address:	Report No MRB2019-10-1419	Date Of Calibration 21/10/2019
Omega M62			Page 27
		Previous Calibration date UNKNOWN	Next Calibration Due 20/10/2020
Description and Identification of the Equipment:-			
I.D. No	9050787	Description T-Wrench	Model Details
Serial No		Manu Teng Tools	operation range 5-25 nms
Cal, Freq,	6 Monthly	Location	Customer Spec +/- 4%

Calibration Details	Traceability Information
TEST 1 - Cust 25 nms TEST 1 - Master 25.4 nms TEST 1 - Result PASS TEST 2 - Cust 15 nms TEST 2 - Actual 15.2 nms TEST 2 - Result PASS TEST 3 - Cust 5 nms TEST 3 - Actual 5 nms TEST 3 - Result PASS	Below are details of the master certificate:- equipment which has been calibrated at a UKAS laboratory to a specification traceable to national standards. Below are details of the master certificate:- ASSET CODE: 17475 EQUIP TYPE: Torque Wrench Analyser MANU: Norbar Torque Tools LTD CAL CERT No: 237273 UKAS LAB No: 0256 LAST CAL: 09/04/2019 NEXT CAL DUE: 09/04/2020
Additional comments/observations:-	CALIBRATION RESULT Pass

Calibration carried out by	Calibration performed by Engineer	Signatory for Calibration & Consultancy Service	
MRB Workshop services Ltd Maryland House 23 Rutland street Swinton Manchester, M27 6AU	Name A.LEES Position Engineer Sign A.LEES	Name M.BATES Date 13/11/2019 Signature	

