

MRB Workshop Services
Maryland House
23 Rutland Street
Swinton
Manchester
M27 6AU
Tel: 0161 793 4040
Fax: 0161 822 0070



Total project
management.....
Total workshop
solution.....

CERTIFICATE OF CALIBRATION

The listed equipment has been tested. Calibrated and has performed in accordance with the CUSTOMERS SPECIFICATION. Each of the readings listed are within tolerance laid down unless indicated otherwise

Customer details: Lanes for drains Monton	Site Address:	Report No MRB2019-10-1463	Date Of Calibration 31/10/2019
			Page 2
		Previous Calibration date UNKNOWN	Next Calibration Due 31/10/2020
Description and Identification of the Equipment:-			
I.D. No 119507	Description T-wrench	Model Details	
Serial No	Manu Britool	operation range 30-151 lbf ft	
Cal, Freq, 6 Monthly	Location	Customer Spec +/- 4%	

Calibration Details	Traceability Information
TEST 1 - Cust 151 lbf ft TEST 1 - Master 153.8 lbf ft TEST 1 - Result PASS TEST 2 - Cust 107 lbf ft TEST 2 - Actual 108.4 lbf ft TEST 2 - Result PASS TEST 3 - Cust 52 psi TEST 3 - Actual TEST 3 - Result PASS	Below are details of the master certificate:- equipment which has been calibrated at a UKAS laboratory to a specification traceable to national standards. Below are details of the master certificate:- ASSET CODE: 17475 EQUIP TYPE: Torque wrench analyser MANU: Norbar torque tools ltd CAL CERT No: 237273 UKAS LAB No: 0256 LAST CAL: 05/04/2019 NEXT CAL DUE: 05/04/2020
Additional comments/observations:-	CALIBRATION RESULT Pass

Calibration carried out by	Calibration performed by Engineer	Signatory for Calibration & Consultancy Service
MRB Workshop services Ltd Maryland House 23 Rutland street Swinton Manchester, M27 6AU	Name A.LEES Position Engineer Sign A.LEES	Name M.BATES Date 18/11/2019 Signature

