

MRB Workshop Services
Maryland House
23 Rutland Street
Swinton
Manchester
M27 6AU
Tel: 0161 793 4040
Fax: 0161 822 0070



Total project
management.....
Total workshop
solution.....

CERTIFICATE OF CALIBRATION

The listed equipment has been tested. Calibrated and has performed in accordance with the CUSTOMERS SPECIFICATION. Each of the readings listed are within tolerance laid down unless indicated otherwise

Customer details: Lanes for drains Monton	Site Address:	Report No	Date Of Calibration
		MRB2019-10-1465	31/10/2019
			Page 4
		Previous Calibration date	Next Calibration Due
		UNKNOWN	31/10/2020
Description and Identification of the Equipment:-			
I.D. No	822247	Description	T-wrench
Serial No		Manu	Sealey
Cal, Freq,	6 Monthly	Location	
		Model Details	operation range 5-82 lbf ft
		Customer Spec	+/- 4%

Calibration Details	Traceability Information
TEST 1 - Cust 82 lbf ft TEST 1 - Master 82.9 lbf ft TEST 1 - Result PASS TEST 2 - Cust 56 lbf ft TEST 2 - Actual 58 lbf ft TEST 2 - Result PASS TEST 3 - Cust 20 lbf ft TEST 3 - Actual 20.3 lbf ft TEST 3 - Result PASS	Below are details of the master certificate:- equipment which has been calibrated at a UKAS laboratory to a specification traceable to national standards. Below are details of the master certificate:- ASSET CODE: 17475 EQUIP TYPE: Torque wrench analyser MANU: Norbar torque tools ltd CAL CERT No: 237273 UKAS LAB No: 0256 LAST CAL: 05/04/2019 NEXT CAL DUE: 05/04/2020
Additional comments/observations:-	CALIBRATION RESULT Pass

Calibration carried out by	Calibration performed by Engineer	Signatory for Calibration & Consultancy Service
MRB Workshop services Ltd	Name A.LEES	Name M.BATES
Maryland House	Position Engineer	Date 18/11/2019
23 Rutland street		
Swinton	Sign A.LEES	Signature
Manchester, M27 6AU		

