

MRB Workshop Services
 Maryland House
 23 Rutland Street
 Swinton
 Manchester
 M27 6AU
 Tel: 0161 793 4040
 Fax: 0161 822 0070



Total project
 management.....
 Total workshop
 solution.....

CERTIFICATE OF CALIBRATION

The listed equipment has been tested. Calibrated and has performed in accordance with the CUSTOMERS SPECIFICATION. Each of the readings listed are within tolerance laid down unless indicated otherwise

Customer details: Halton Council Widnes	Site Address:	Report No	Date Of Calibration
		MRB2019-10-1475	11/10/2019
			Page 6
		Previous Calibration date	Next Calibration Due
		UNKNOWN	11/10/2019

Description and Identification of the Equipment:-

I.D. No	PL002	Description	T-Wrench	Model Details	
Serial No		Manu	Britool	operation range	20-100 lbf ft
Cal, Freq,	6 Monthly	Location		Customer Spec	+/- 4%

Calibration Details	Traceability Information
TEST 1 - Cust 100 lbf ft TEST 1 - Master 98.7 lbf ft TEST 1 - Result PASS TEST 2 - Cust 60 lbf ft TEST 2 - Actual 59.4 lbf ft TEST 2 - Result PASS TEST 3 - Cust 30 lbf ft TEST 3 - Actual 30.1 lbf ft TEST 3 - Result PASS	Below are details of the master certificate:- equipment which has been calibrated at a UKAS laboratory to a specification traceable to national standards. Below are details of the master certificate:- ASSET CODE: 17475 EQUIP TYPE: Torque wrench analyser MANU: Norbar torque tools ltd CAL CERT No: 237273 UKAS LAB No: 0256 LAST CAL: 05/04/2019 NEXT CAL DUE: 05/04/2020

Additional comments/observations:-	CALIBRATION RESULT Pass
------------------------------------	-----------------------------------

Calibration carried out by	Calibration performed by Engineer	Signatory for Calibration & Consultancy Service
MRB Workshop services Ltd Maryland House 23 Rutland street Swinton Manchester, M27 6AU	Name A.LEES Position Engineer Sign A.LEES	Name M.BATES Date 18/11/2019 Signature

