

MRB Workshop Services
 Maryland House
 23 Rutland Street
 Swinton
 Manchester
 M27 6AU
 Tel: 0161 793 4040
 Fax: 0161 822 0070



Total project
 management.....
 Total workshop
 solution.....

CERTIFICATE OF CALIBRATION

The listed equipment has been tested. Calibrated and has performed in accordance with the CUSTOMERS SPECIFICATION. Each of the readings listed are within tolerance laid down unless indicated otherwise

Customer details: Halton Council	Site Address:	Report No MRB2019-10-1483	Date Of Calibration 11/10/2019
Widnes			Page 14
		Previous Calibration date UNKNOWN	Next Calibration Due 11/10/2019
Description and Identification of the Equipment:-			
I.D. No	PG002	Description	Pressure gauge
Serial No		Manu	Bendix
Cal, Freq,	6 Monthly	Location	Customer Spec
		Model Details	operation range 0-20 Bar
			Customer Spec +/- 4%

Calibration Details	Traceability Information
TEST 1 - Cust 7 Bar TEST 1 - Master 6.9 Bar TEST 1 - Result PASS TEST 2 - Cust 14 Bar TEST 2 - Actual 14 Bar TEST 2 - Result PASS TEST 3 - Cust 20 Bar TEST 3 - Actual 20 Bar TEST 3 - Result PASS	Below are details of the master certificate:- equipment which has been calibrated at a UKAS laboratory to a specification traceable to national standards. Below are details of the master certificate:- ASSET CODE: 17475 EQUIP TYPE: Torque wrench analyser MANU: Norbar torque tools ltd CAL CERT No: 237273 UKAS LAB No: 0256 LAST CAL: 05/04/2019 NEXT CAL DUE: 05/04/2020
Additional comments/observations:-	CALIBRATION RESULT Pass

Calibration carried out by	Calibration performed by Engineer	Signatory for Calibration & Consultancy Service
MRB Workshop services Ltd Maryland House 23 Rutland street Swinton Manchester, M27 6AU	Name A.LEES Position Engineer Sign A.LEES	Name M.BATES Date 18/11/2019 Signature

