

MRB Workshop Services
 Maryland House
 23 Rutland Street
 Swinton
 Manchester
 M27 6AU
 Tel: 0161 793 4040
 Fax: 0161 822 0070



Total project
 management.....
 Total workshop
 solution.....

CERTIFICATE OF CALIBRATION

The listed equipment has been tested. Calibrated and has performed in accordance with the CUSTOMERS SPECIFICATION. Each of the readings listed are within tolerance laid down unless indicated otherwise

Customer details: Sefton Council Liverpool	Site Address:	Report No	Date Of Calibration
		MRB2019-10-1485	12/11/2019
			Page 5
		Previous Calibration date	Next Calibration Due
		UNKNOWN	12/11/2020

Description and Identification of the Equipment:-

I.D. No	191018	Description	T-Wrench
Serial No		Manu	teng tool
Cal, Freq,	6 Monthly	Location	
		Model Details	operation range 51-258 lbf ft
		Customer Spec	+/- 4%

Calibration Details	Traceability Information
TEST 1 - Cust 258 lbf ft TEST 1 - Master 259.8 lbf ft TEST 1 - Result PASS TEST 2 - Cust 154 lbf ft TEST 2 - Actual 154.9 lbf ft TEST 2 - Result PASS TEST 3 - Cust 51 lbf ft TEST 3 - Actual 51.3 lbf ft TEST 3 - Result PASS	Below are details of the master certificate:- equipment which has been calibrated at a UKAS laboratory to a specification traceable to national standards. Below are details of the master certificate:- ASSET CODE: 17475 EQUIP TYPE: Torque Wrench analyser MANU: Norbar Torque tools LTD CAL CERT No: 237273 UKAS LAB No: 0256 LAST CAL: 05/04/2019 NEXT CAL DUE: 05/04/2020

Additional comments/observations:-	CALIBRATION RESULT Pass
------------------------------------	-----------------------------------

Calibration carried out by	Calibration performed by Engineer	Signatory for Calibration & Consultancy Service
MRB Workshop services Ltd Maryland House 23 Rutland street Swinton Manchester, M27 6AU	Name A.LEES Position Engineer Sign A.LEES	Name M.BATES Date 18/11/2019 Signature

