

MRB Workshop Services  
 Maryland House  
 23 Rutland Street  
 Swinton  
 Manchester  
 M27 6AU  
 Tel: 0161 793 4040  
 Fax: 0161 822 0070



Total project  
 management.....  
 Total workshop  
 solution.....

## CERTIFICATE OF CALIBRATION

The listed equipment has been tested. Calibrated and has performed in accordance with the CUSTOMERS SPECIFICATION. Each of the readings listed are within tolerance laid down unless indicated otherwise

<b>Customer details:</b> Hyndburn Council Accrington	<b>Site Address:</b>	<b>Report No</b>	<b>Date Of Calibration</b>
		MRB2019-10-1474	06/11/2019
			<b>Page</b> 2
		<b>Previous Calibration date</b>	<b>Next Calibration Due</b>
		UNKNOWN	06/11/2020

Description and Identification of the Equipment:-

<b>I.D. No</b>	2263187	<b>Description</b>	T-Inflator
<b>Serial No</b>		<b>Manu</b>	PCL
<b>Cal, Freq,</b>	6 Monthly	<b>Location</b>	
		<b>Model Details</b>	operation range 10-170 psi
		<b>Customer Spec</b>	+/- 4%

<b>Calibration Details</b>	<b>Traceability Information</b>
TEST 1 - Cust 20 psi TEST 1 - Master 21 psi TEST 1 - Result PASS  TEST 2 - Cust 70 psi TEST 2 - Actual 71 psi TEST 2 - Result PASS  TEST 3 - Cust 130 psi TEST 3 - Actual 131 psi TEST 3 - Result PASS	Below are details of the master certificate:- equipment which has been calibrated at a UKAS laboratory to a specification traceable to national standards. Below are details of the master certificate:-  ASSET CODE: 8824 EQUIP TYPE: Tyre Pressure gauge tester MANU: PCL CAL CERT No: 549003 UKAS LAB No: 0495 LAST CAL: 05/04/2019 NEXT CAL DUE: 05/04/2020

Additional comments/observations:-	<b>CALIBRATION RESULT</b> Pass
------------------------------------	-----------------------------------

<b>Calibration carried out by</b>	<b>Calibration performed by Engineer</b>	<b>Signatory for Calibration &amp; Consultancy Service</b>
MRB Workshop services Ltd Maryland House 23 Rutland street Swinton Manchester, M27 6AU	Name A.LEES Position Engineer  Sign A.LEES	Name M.BATES Date 18/11/2019  Signature

